

<b>Notice of Allowability</b>	<b>Application No.</b>	<b>Applicant(s)</b>	
	10/089,680	MITANI ET AL.	
	<b>Examiner</b>	<b>Art Unit</b>	
	Brian J. Sines	1743	

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--**

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to the response filed 6/8/2006.
2. ☒ The allowed claim(s) is/are 1-20.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☒ All    b) ☐ Some\*    c) ☐ None    of the:
  1. ☒ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
  5. ☐ CORRECTED DRAWINGS ( as "replacement sheets") must be submitted.
    - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review ( PTO-948) attached
      - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
    - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

**Attachment(s)**

- |  |  |
|--|--|
| 1. <input type="checkbox"/> Notice of References Cited (PTO-892)   | 5. <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)            |
| 2. <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)   | 6. <input type="checkbox"/> Interview Summary (PTO-413),<br>Paper No./Mail Date _____. |
| 3. <input checked="" type="checkbox"/> Information Disclosure Statements (PTO-1449 or PTO/SB/08),<br>Paper No./Mail Date _____ | 7. <input type="checkbox"/> Examiner's Amendment/Comment                               |
| 4. <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit<br>of Biological Material                     | 8. <input checked="" type="checkbox"/> Examiner's Statement of Reasons for Allowance   |
|  | 9. <input type="checkbox"/> Other _____.   |

## **DETAILED ACTION**

### ***Allowable Subject Matter***

Claims 1 – 20 are allowed.

The following is an examiner's statement of reasons for allowance:

Regarding claim 1, the cited prior art neither teach nor fairly suggest a method for controlling a production process without having to resort to the preparation of a calibration curve for predicting data for analysis samples, the method comprising:

taking an absorbance spectrum for each of a plurality of standard samples

collected from a production process step in an analysis range including the near-infrared region;

constructing a data base from a differentiation curve of a near-infrared

spectrum chart obtained using a plurality of production products that had been judged by conventional chemical analysis to be rated products, by calculating standard deviations and the average intensity of the standard samples (standard average intensity) in respect of each of the wavelengths selected from the spectrum included in the analysis range at a constant interval;

taking an absorbance spectrum in the analysis range for each analysis

sample collected from the production process step and comparing the resulting absorbance spectrum with the data base;

estimating the deviation (analysis deviation) of the intensity of the absorbance

spectrum of each of the analysis samples (analysis intensity) at each of the selected wavelengths from the standard average intensity;

comparing, when the absorbance spectrum includes wavelength(s) at which the analysis deviation of the absorbance of the analysis sample is outside a tolerance limit determined based on the standard deviation, the wavelength showing the analysis deviation of the absorbance outside the tolerance limit with production information given preliminarily in the data base in order to find out one or more control factors responsive to the analysis deviation of absorbance of the analysis sample;

estimating control data for reclaiming the production process based on the one or more control factors; and

controlling the production process so as to obtain production product within the tolerance limit by inputting the control data to the production process step.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Brian J. Sines whose telephone number is (571) 272-1263. The examiner can normally be reached on Monday - Friday (11 AM - 8 PM EST).

Art Unit: 1743

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill A. Warden can be reached on (571) 272-1267. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

A handwritten signature in black ink, appearing to read "Brian Smith". The signature is written in a cursive, flowing style with a large loop at the end of the last name.